


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0428 Accredited to ISO/IEC 17025:2005	Precisa Limited	
	Issue No: 018	Issue date: 04 November 2016
	12 Cromwell Business Centre Howard Way Newport Pagnell Milton Keynes Buckinghamshire MK16 9QS	Contact: Ms Tina Mason Tel: +44 (0)1908-211900 Fax: +44 (0)1908-211909 E-Mail: info@precisa.co.uk Website: www.precisa.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 12 Cromwell Business Centre Howard Way Newport Pagnell Milton Keynes Buckinghamshire MK16 9QS	Local contact Tina Mason	Calibration of Non Automatic Weighing Machine P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At Customers Premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Calibration Engineer Calibration of Non Automatic Weighing Machine	S



0428
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Precisa Limited
Issue No: 018 Issue date: 04 November 2016

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
NON-AUTOMATIC WEIGHING MACHINES See notes 1 and 2.	200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 40 kg 50 kg 100 kg 200 kg 400 kg	0.018 mg 0.02 mg 0.025 mg 0.033 mg 0.043 mg 0.05 mg 0.068 mg 0.083 mg 0.13 mg 0.25 mg 0.63 mg 1.26 mg 3.83 mg 7.98 mg 0.019 g 0.039 g 0.10 g 0.16 g 1.7 g 4.4 g 8.8 g	1. Weights are available in OIML class: E2 1 mg to 500 g. max grouped load 1 kg F1 10 mg to 20 kg. max grouped load 46 kg. F2 2 x 10 kg. maxgrouped load 66 kg (with F1 masses). M1 5 kg to 20 kg. max grouped load 400 kg. 2. Other loads within the overall listed range may also be used.	S
END				